

**LOT**

**CX 513**

Rev 1

**CAL**



**(Exp.) 2026-01-05**  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	
		EMI CRP	LC-767G	MS CRP	XLR	DX NEXUS DF NEXUS	MS60		60 60 C+ ES60	XL80	
GB WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>9.0</b>	<b>9.0</b>	<b>9.0</b>	<b>8.9</b>	<b>9.3</b>	<b>8.8</b>		<b>8.9</b>	<b>8.9</b>	± 0.2
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>4.31</b>	<b>4.32</b>	<b>4.28</b>	<b>4.33</b>	<b>4.36</b>	<b>4.27</b>		<b>4.35</b>	<b>4.33</b>	± 0.06
HB HGB	g/dl	<b>12.8</b>	<b>12.8</b>	<b>12.8</b>	<b>13.0</b>	<b>12.9</b>	<b>12.8</b>		<b>13.0</b>	<b>13.0</b>	± 0.2
	g/l	<b>128</b>	<b>128</b>	<b>128</b>	<b>130</b>	<b>129</b>	<b>128</b>		<b>130</b>	<b>130</b>	± 2
	mmol/l	<b>7.95</b>	<b>7.95</b>	<b>7.95</b>	<b>8.07</b>	<b>8.01</b>	<b>7.95</b>		<b>8.07</b>	<b>8.07</b>	± 0.12
HT HCT	%	<b>36.2</b>	<b>36.1</b>	<b>34.5</b>	<b>35.5</b>	<b>35.1</b>	<b>34.4</b>		<b>35.7</b>	<b>35.5</b>	± 1.0
	l/l	<b>0.362</b>	<b>0.361</b>	<b>0.345</b>	<b>0.355</b>	<b>0.351</b>	<b>0.344</b>		<b>0.357</b>	<b>0.355</b>	± 0.010
PLA PLT	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>234</b>	<b>236</b>	<b>233</b>	<b>232</b>	<b>233</b>	<b>233</b>		<b>238</b>	<b>232</b>	± 10
VMP MPV	µm <sup>3</sup> ; fl	<b>8.4</b>	<b>8.3</b>	N/A	N/A	N/A	N/A		N/A	N/A	± 0.5

Ref: TEMP-0387 Rev.53