

REF 0401005

REAGENT 0.5 L

IVD 

HORIBA ABX SAS
Parc Euromédecine
Rue du Caducée
BP 7290
34184 Montpellier Cedex 4
FRANCE

ABX Minoclair 0.5L

- Advia 60
- ABX Micros 60 / ABC Vet
- ABX Micros ES60 / ESV60
- ABX Micros CRP200
- ABX Pentra 60 / 60C+
- ABX Pentra XL80
- Pentra XLR
- ABX Pentra DX120 / DF120
- Pentra DX Nexus / DF Nexus
- Pentra ES60 / MS60 / MS CRP
- Micros Care ST
- Microsemi CRP
- Yumizen H500 OT / CT / H550
- Yumizen H1500 / H2500

Hematology Devices (for *in vitro* diagnostic use)

Intended Use ^{a b}

ABX Minoclair 0.5L is a chemical solution intended for *in vitro* diagnostic use and designed for the cleaning of HORIBA Medical blood cell counters. Clinical laboratories use.

Warnings and Precautions ^c

- **ABX Minoclair 0.5L** is for professional *in vitro* diagnostic use only. For laboratory use.
- It is the user's responsibility to verify that this document is applicable to the product use.
- This reagent is classified as hazardous in compliance with regulation (EC) N°.1272/2008.
- **Warning**
H290: May be corrosive to metals.
EUH031: Contact with acids liberates toxic gas.
P234: Keep only in original container.
P390: Absorb spillage to prevent material damage.
P406: Store in corrosive resistant container with a resistant inner liner.
- Users are advised to wear approved protective clothing when handling chemical products: lab coat, gloves, and eye protection.
- Observe the standard laboratory precautions for use and follow national or local health and safety guidelines.
- User must be trained by a HORIBA Medical representative before attempting to operate the device.
- In the event of a malaise following skin contact, ingestion, or inhalation, consult a doctor.
- Please refer to the Safety Data Sheet (SDS) associated with **ABX Minoclair 0.5L**.

- Any serious incident that has occurred in relation to the device shall be reported to the manufacturer and the competent authority of the country in which the user and/or the patient is established.
- The reagent containers are disposable and should be disposed of in accordance with the local legal requirements.
- For technical assistance, you can call +33 (0)4 67 14 15 16.
- This reagent is destined for use with HORIBA Medical blood cell counters specified above. HORIBA Medical cannot guarantee the correct functioning of this reagent with instruments other than those specified above, or with instruments not manufactured by HORIBA Medical.

Waste Management

Please refer to local legal requirements.

Microbiological State

Not applicable.

Description and Composition

Description:

Basic, limpid and yellowish aqueous solution. Smells of bleach.

^aModification: new reagent leaflet form.

^bModification: modification of intended use.

^cModification: recommendation added.

ABX Minocclair 0.5L

Composition:

Chemical cleaning agent	< 5%
Stabilizer	< 1%

Storage and Stability

- **Storage condition (before opening):** 18-25°C (65-77°F).
Do not freeze.
- **Open stability:** 6 months maximum at 18-25°C (65-77°F) after opening and within the expiration limit.
- **Expiration date:** refer to "expiration date" reagent packaging label.

Materials Required but not Provided

- Automated hematology analyzer.
- Calibrator: **ABX Minocal**.
- Control: refer to the user manual for the specific control used with your instrument.
- Standard laboratory equipment.

Specimen

Sample collection:

All blood samples should be collected using proper technique! Consider all specimens, reagents, calibrators, controls, etc. that contain human specimen extracts as potentially infectious and follow biosafety practices (1, 2). When collecting blood specimens, venous blood is recommended, but arterial blood may also be used in extreme cases. Blood collection must be placed in vacuum or atmospheric collection tubes (3, 4). The sample collection tube has to be filled to the exact quantity of blood indicated on the tube itself to avoid variations in the results.

Recommended anti-coagulant:

The recommended anticoagulant is K₃-EDTA with the proper proportion of blood to anticoagulant as specified by the tube manufacturer. K₂-EDTA is an acceptable alternative, as long as the sample collection is made in normal conditions. Otherwise, blood clots may be possible.

Blood sample stability:

Sample stability at low temperature: Ten "normal" and ten "pathological" specimens were collected from the routine

laboratory workload and stored at 4°C. Sample stability was assessed over a period of 72 hours. The results (mean of ten tests) conclude with a relative sample stability claim of:

- 48 hours for the CBC parameters
- 24 hours for the DIFF parameters

Sample stability at room temperature: Ten "normal" and ten "pathological" specimens were collected from the routine laboratory workload and stored at room temperature (25°C). Sample stability was assessed over a period of 72 hours. The results (mean of ten tests) conclude with a relative sample stability claim of:

- 48 hours for the CBC parameters
- 24 hours for the DIFF parameters

Microsampling:

Instrument sampling mode enables the user to work with microsamples for pediatrics and geriatrics (refer to the instrument user manual for the minimum blood sample volume). These microsamples can only be used in the following conditions:

- The tube must always be held in vertical position.
- Blood mixing must be obtained by slight tapping on the tube. Do not rotate the tube for mixing, otherwise the blood will be spread on the tube side, and the minimum required level will be lost.

Mixing:

Blood samples must be gently and thoroughly mixed just before sampling. This ensures a homogeneous mixture for measurement.

Procedure

This reagent is ready to use.

ABX Minocclair 0.5L is used during auto-concentrated cleaning procedure. Refer to the instrument user manual for running auto-concentrated cleaning.

Methodology

ABX Minocclair 0.5L has an oxydizing action to clean up the hydraulic parts of the instrument.

ABX Minoclair 0.5L

Performance Characteristics and Limitations of the Method

Refer to the user manual for the performance characteristics of the instrument and the limitations of the analyses on instrument parameters.

Calculation and Interpretation of Analytical Results

Refer to the instrument user manual for calculation and interpretation of analytical results.

Changes in the Procedure and in the Performance

Packaging spoiling

In case of protective packaging spoiling, do not use **ABX Minoclair 0.5L** if the damages might have an effect on the product performance.

Signs of deterioration

In the event of any signs of physical or chemical deterioration (turbidity, change in colour etc.) **ABX Minoclair 0.5L** should be replaced.

Temperature limits

Do not use **ABX Minoclair 0.5L** if it has been frozen or kept at excessive heat.

Before using **ABX Minoclair 0.5L**, make sure it has reached the operating temperature conditions as described in the instrument user manual.

Internal Quality Control

HORIBA Medical control bloods must be used to periodically assess the integrity of the reagents and the instrument in the specified ranges.

HORIBA Medical offers an Online Interlaboratory Comparison Program (QCP) which provides internet access to:

- Submit Internal Quality Control results online.
- Monitor analytical performances and compare directly with hundreds of laboratories worldwide.
- Obtain real time peer group statistical reports from QCP

More informations are available at:

<http://qcp.horiba-abx.com>

Traceability of Calibrators and Control Materials

Not applicable.

Reference Intervals

Not applicable.

Reference

1. Occupational Safety and Health Standards: bloodborne pathogens. (29 CFR 1910. 1030). Federal Register July 1, 1998; **6**: 267-280.
2. Protection of Laboratory Workers From Occupationally Acquired Infections; Approved Guideline - Fourth Edition. CLSI (NCCLS), document M29-A4 (2014) **34** (18).
3. Procedures for the Collection of Diagnostic Blood Specimens by Venipuncture; Approved Standard - Sixth Edition. CLSI (NCCLS), document H3-A6 (2007) **27** (26).
4. Procedures and Devices for the Collection of Diagnostic Capillary Blood Specimens; Approved Standard - Sixth Edition. CLSI (NCCLS), document H4-A6 (2008) **28** (25).

