

Hematology Devices (for in vitro diagnostic use)

ABX Minoclair

Exclusive use:


ABX Micros ABC Vet
ABX Micros CRP
ABX Micros CRP200
ABX Micros 60
ABX Pentra 60
ABX Pentra 60 C+
ABX Pentra 80
ABX Pentra XL 80
ABX Pentra 120
ABX Pentra 120 Retic
ABX Pentra DX 120
ABX Slide Preparation System

21/08/07
A95A00029DEN

REF 0401005 (0,5L)

REAGENT 0,5L

IVD 

 **HORIBA ABX**
BP 7290 - 34184
Montpellier
cedex 4 - France

1. Functions

Cleaning and bleaching solution for HORIBA ABX blood cell counters.

Measurement procedure to be followed in using the device:

Principle of the method, specific analytical performance characteristics : see «Section: Specifications» in the instrument User Manual.

2. Conservation & Expiration

Storage conditions: Stored at 18°C (65°F) to 25°C (77°F).

Expiration date: refer to «expiration date» reagent packaging label.

3. Measurements, principles & results

Directions for use: see «Section: Reagent Location and connection» in the instrument User Manual.

Measuring Principles : see «Section: Technology» in the instrument User Manual.

Results: Refer to the instrument User Manual

Performance data: see «Section: Specifications» in the instrument User Manual.

Note: if performance changes, call your HORIBA ABX representative.

4. Composition & Handling precautions

Composition:

Sodium hypochlorite: 10,5 % at 13% of active chlorine ^a.

Sodium hydroxide: 0,26%

a.From index C to index D : modification of the composition description

pH: 12,4 +/- 0,5 (T = 20°C)

Resistivity: Not available

Description: Yellowish liquid.

Handling Precautions: Avoid contact with eyes, skin and clothing. Wear laboratory gloves when handling the product. The product may be harmful if ingested or inhaled. Keep the bottle closed when not in use. Please refer to the MSDS associated with the reagent.

Specimen Collection and Mixing: see «Section: Specimen collection and Mixing» in the instrument User Manual.

5. Limitations & waste disposal

Limitations: see «Section: Specifications» in the instrument User Manual.

Safe Waste Disposal: Follow your laboratory's protocol when neutralizing and disposing of waste. Please refer to the MSDS associated with the reagent.

6. Concentrated cleaning exclusive use

ABX Minoclair may also be used as disinfectant and cleaning product on HORIBA ABX Blood cell counters: see «Section: Concentrated cleaning» in the user manual, on the following instruments: ABX Micros ABC Vet, ABX Micros 60, ABX Micros CRP, ABX Micros CRP200, ABX Pentra 60, ABX Pentra 60C+, ABX Pentra 80, ABX Pentra XL80.

