

# Hematology Devices (for in vitro diagnostic use)

## ABX Minoclair

### Exclusive use:


ABX Micros ABC Vet  
ABX Micros CRP  
ABX Micros CRP200  
ABX Micros 60  
ABX Pentra 60  
ABX Pentra 60 C+  
ABX Pentra 80  
ABX Pentra XL 80  
ABX Pentra 120/120 Retic  
ABX Pentra DX 120  
ABX Pentra DF 120<sup>a</sup>

28/03/08  
A95A00029EEN

**REF** 0401005 (0,5L)

**REAGENT** 0,5L

**IVD** 

 **HORIBA ABX**  
BP 7290 - 34184  
Montpellier  
cedex 4 - France

## 1. Functions

Cleaning and bleaching solution for HORIBA ABX blood cell counters.

### Measurement procedure to be followed in using the device:

Principle of the method, specific analytical performance characteristics : see «Section: Specifications» in the instrument User Manual.

## 2. Conservation & Expiration

**Storage conditions:** Stored at 18°C (65°F) to 25°C (77°F).

**Open stability:** maximum 6 months after a vial has been opened<sup>b</sup>.

**Expiration date:** refer to «expiration date» reagent packaging label.

## 3. Measurements, principles & results

**Directions for use<sup>c</sup>:** see «Section: Maintenance & Troubleshooting» in the instrument User Manual.

**Measuring Principles :** see «Section: Technology» in the instrument User Manual.

**Results:** Refer to the instrument User Manual

**Performance data:** see «Section: Specifications» in the instrument User Manual.

Note: if performance changes, call your HORIBA ABX representative.

## 4. Composition & Handling precautions

### Composition:

Sodium hypochlorite:.....10,5 % at 13% of active chlorine.

Sodium hydroxide:.....0,26%

**pH:** 12,4 +/- 0,5 (T = 20°C)

**Resistivity:** Not available

**Description:** Yellowish liquid.

**Handling Precautions:** Avoid contact with eyes, skin and clothing. Wear laboratory gloves when handling the product. The product may be harmful if ingested or inhaled. Keep the bottle closed when not in use. Please refer to the MSDS associated with the reagent.

**Specimen Collection and Mixing:** see «Section: Specimen collection and Mixing» in the instrument User Manual.

## 5. Limitations & waste disposal

**Limitations:** see «Section: Specifications» in the instrument User Manual.

**Safe Waste Disposal:** Follow your laboratory's protocol when neutralizing and disposing of waste. Please refer to the MSDS associated with the reagent.

## 6. Concentrated cleaning exclusive use

ABX Minoclair may also be used as disinfectant and cleaning product on HORIBA ABX Blood cell counters: see «Section: Concentrated cleaning» in the user manual, on the following

a.Modification from index D to E: ABX Pentra DF 120

b.Modification from index D to E: Open stability

c.Modification from index D to E: Modification (Directions for use)

instruments: ABX Micros ABC Vet, ABX Micros 60, ABX Micros CRP, ABX Micros CRP200, ABX Pentra 60, ABX Pentra 60C+, ABX Pentra 80, ABX Pentra XL 80, ABX Pentra 120, ABX Pentra 120 Retic, ABX Pentra DX 120, ABX Pentra DF 120.