

Hematology Devices (for in vitro diagnostic use)

ABX Minoclair

Exclusive use:

ABX Micros ABC Vet
ABX Micros CRP/CRP200
ABX Micros 60
ABX Micros ES60/ESV60
ABX Pentra 60
ABX Pentra 60 C+
ABX Pentra 80
ABX Pentra XL 80
ABX Pentra 120
ABX Pentra 120 Retic
ABX Pentra DX 120/DF 120

04/02/09
A95A00029FEN

REF 0401005 (0,5L)

REAGENT 0,5L

IVD 



HORIBA ABX SAS
BP 7290 - 34184
Montpellier
cedex 4 - France

1. Functions

Cleaning and bleaching solution for HORIBA Medical blood cell counters^a.

Measurement procedure to be followed in using the device:

Principle of the method, specific analytical performance characteristics : see «Section: Specifications» in the instrument User Manual.

2. Conservation & Expiration

Storage conditions: Stored at 18°C (65°F) to 25°C (77°F).

Open stability: maximum 6 months after a vial has been opened.

Expiration date: refer to «expiration date» reagent packaging label.

3. Measurements, principles & results

Directions for use: see «Section: Maintenance & Troubleshooting» in the instrument User Manual.

Measuring Principles : see «Section: Technology» in the instrument User Manual.

Results: Refer to the instrument User Manual

Performance data: see «Section: Specifications» in the instrument User Manual.

Note: if performance changes, call your HORIBA Medical representative.

4. Composition & Handling precautions

Composition:

Sodium hypochlorite:.....10,5 % at 13% of active chlorine.

Sodium hydroxide:.....0,26%

pH: 12,4 +/- 0,5 (T = 20°C)

Resistivity: Not available

Description: Yellowish liquid.

Handling Precautions: Avoid contact with eyes, skin and clothing. Wear laboratory gloves when handling the product. The product may be harmful if ingested or inhaled. Keep the bottle closed when not in use. Please refer to the MSDS associated with the reagent.

Specimen Collection and Mixing: see «Section: Specimen collection and Mixing» in the instrument User Manual.

5. Limitations & waste disposal

Limitations: see «Section: Specifications» in the instrument User Manual.

Safe Waste Disposal: Follow your laboratory's protocol when neutralizing and disposing of waste. Please refer to the MSDS associated with the reagent.

6. Concentrated cleaning exclusive use

ABX Minoclair may also be used as disinfectant and cleaning product on HORIBA Medical Blood cell counters: see «Section: Concentrated cleaning» in the user manual, on the following

a.Modification from index E to F: ABX Micros ES60/ESV60

instruments: ABX Micros ABC Vet, ABX Micros 60, ABX Micros ES60, ABX Micros ESV60, ABX Micros CRP, ABX Micros CRP200, ABX Pentra 60, ABX Pentra 60C+, ABX Pentra 80, ABX Pentra XL 80, ABX Pentra 120, ABX Pentra 120 Retic, ABX Pentra DX 120, ABX Pentra DF 120.