

# Hematology Devices (for in vitro diagnostic use)

## ABX Minocclair

### Exclusive use:

ABX Micros ABC Vet  
ABX Micros CRP/CRP200  
ABX Micros 60  
ABX Micros ES60/ESV60  
ABX Pentra 60  
ABX Pentra 60 C+  
Pentra ES60  
ABX Pentra 80  
ABX Pentra XL 80  
ABX Pentra 120  
ABX Pentra 120 Retic  
ABX Pentra DX 120/DF 120  
ADVIA 60

29/03/10  
A95A00029HEN

**REF** 0401005 (0,5L)

**REAGENT** 0,5L

**IVD** 



**HORIBA ABX SAS**  
BP 7290 - 34184  
Montpellier  
cedex 4 - France

## 1. Functions

Cleaning and bleaching solution for blood cell counters<sup>a</sup>.

### Measurement procedure to be followed in using the device:

Principle of the method, specific analytical performance characteristics, see «Section: Specifications» in the instrument User Manual.

## 2. Conservation & Expiration

**Storage conditions:** Stored at 18°C (65°F) to 25°C (77°F).

**Open stability:** maximum 6 months after a vial has been opened.

**Expiration date:** refer to «expiration date» reagent packaging label.

## 3. Measurements, principles & results

**Directions for use:** see «Section: Maintenance & Troubleshooting» in the instrument User Manual.

**Measuring Principles :** see «Section: Technology» in the instrument User Manual.

**Results:** Refer to the instrument User Manual

**Performance data:** see «Section: Specifications» in the instrument User Manual.

Note: if performance changes, call your HORIBA Medical representative.

## 4. Composition & Handling precautions

### Composition:

Sodium hypochlorite:..... 11,4 % at 13% of active chlorine.

Sodium hydroxide:..... 0,35%

a.Modification from index G to H: ADVIA 60 added.

**pH:** 12,5 +/- 0,5 (T = 20°C)

**Resistivity:** Not available

**Description:** Yellowish liquid.

### Handling Precautions:

C - Corrosive

R35 : causes severe burns.

S26 : In case of contact with eyes, rinse immediately with plenty of water and seek medical advice.

S36/37/39 : Wear suitable protective clothing, gloves and eye/face protection.

S45 : In case of accident or if you feel unwell, seek medical advice immediately (show label where possible).

S60 : This material and/or its container must be disposed of as hazardous waste.

Avoid contact with eyes, skin and clothing. Wear laboratory gloves when handling the product. The product may be harmful if ingested or inhaled. Keep the bottle closed when not in use. Please refer to the MSDS associated with the reagent.

## 5. Limitations & waste disposal

**Limitations:** see «Section: Specifications» in the instrument User Manual.

**Safe Waste Disposal:** Follow your laboratory's protocol when neutralizing and disposing of waste. Please refer to the MSDS associated with the reagent.

## 6. Concentrated cleaning exclusive use

ABX Minoclair may also be used as disinfectant and cleaning product on blood cell counters: see «Section: Concentrated cleaning» in the user manual, on the following instruments: ABX Micros ABC Vet, ABX Micros 60, ABX Micros ES60, ABX Micros ESV60, ABX Micros CRP, ABX Micros CRP200, ABX Pentra 60, ABX Pentra 60C+, Pentra ES60, ABX Pentra 80, ABX Pentra XL 80, ABX Pentra 120, ABX Pentra 120 Retic, ABX Pentra DX 120, ABX Pentra DF 120, ADVIA 60<sup>a</sup>.

---

a.Modification from index G to H: ADVIA 60 added.