

Hematology Devices (for in vitro diagnostic use)

ABX Minocclair

15/04/2010
A95A00029IEN

Exclusive use:

ABX Micros ABC Vet
ABX Micros CRP/CRP200
ABX Micros 60
ABX Micros ES60/ESV60
ABX Pentra 60
ABX Pentra 60 C+
Pentra ES60
Pentra MS60
ABX Pentra 80
ABX Pentra XL 80
ABX Pentra 120
ABX Pentra 120 Retic
ABX Pentra DX 120/DF 120
ADVIA 60

REF 0401005 (0,5L)

REAGENT 0,5L

IVD 



HORIBA ABX SAS
Parc Euromédecine
Rue du Caducée
BP 7290
34184 Montpellier Cedex 4
FRANCE

1. Functions

Cleaning and bleaching solution for blood cell counters^a.

Measurement procedure to be followed in using the device:

Principle of the method, specific analytical performance characteristics, see «Section: Specifications» in the instrument User Manual.

2. Conservation & Expiration

Storage conditions: Stored at 18°C (65°F) to 25°C (77°F).

Open stability: maximum 6 months after a vial has been opened.

Expiration date: refer to «expiration date» reagent packaging label.

3. Measurements, principles & results

Directions for use: see «Section: Maintenance & Troubleshooting» in the instrument User Manual.

Measuring Principles : see «Section: Technology» in the instrument User Manual.

Results: Refer to the instrument User Manual

Performance data: see «Section: Specifications» in the instrument User Manual.

Note: if performance changes, call your HORIBA Medical representative.

4. Composition & Handling precautions

Composition:

Sodium hypochlorite:..... 11,4 % at 13% of active chlorine.

Sodium hydroxide:..... 0,35%

a.Modification from index H to I: Pentra MS60 added.

pH: 12,5 +/- 0,5 (T = 20°C)

Resistivity: Not available

Description: Yellowish liquid.

Handling Precautions:

C - Corrosive

R35 : causes severe burns.

S26 : In case of contact with eyes, rinse immediately with plenty of water and seek medical advice.

S36/37/39 : Wear suitable protective clothing, gloves and eye/face protection.

S45 : In case of accident or if you feel unwell, seek medical advice immediately (show label where possible).

S60 : This material and/or its container must be disposed of as hazardous waste.

Avoid contact with eyes, skin and clothing. Wear laboratory gloves when handling the product. The product may be harmful if ingested or inhaled. Keep the bottle closed when not in use. Please refer to the MSDS associated with the reagent.

5. Limitations & waste disposal

Limitations: see «Section: Specifications» in the instrument User Manual.

Safe Waste Disposal: Follow your laboratory's protocol when neutralizing and disposing of waste. Please refer to the MSDS associated with the reagent.

6. Concentrated cleaning exclusive use

ABX Minoclair may also be used as disinfectant and cleaning product on blood cell counters: see «Section: Concentrated cleaning» in the user manual, on the following instruments: ABX Micros ABC Vet, ABX Micros 60, ABX Micros ES60, ABX Micros ESV60, ABX Micros CRP, ABX Micros CRP200, ABX Pentra 60, ABX Pentra 60C+, Pentra ES60, ABX Pentra 80, ABX Pentra XL 80, ABX Pentra 120, ABX Pentra 120 Retic, ABX Pentra DX 120, ABX Pentra DF 120, ADVIA 60, Pentra MS60^a.

a.Modification from index H to I: Pentra MS60 added.