




Certificate of Traceability

Manufacturer Certification


This certifies that **HORIBA Medical**, using **Pointe Hemoglobin A1c - HbA1c% - Pentra C200** has participated in and successfully completed the NGSP certification for manufacturers and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: **European Reference Laboratory ESRL#13**

The system evaluated was:


Instrument: Pentra C200	Calibrator Lot: 419301	Application: $Y=(1.15 \times X)-0.53$
Reagent Lot: 427101	Calibrator Assigned Values: 0%, 4.4%, 8.2%, 11.1%, 15.9%	

Date of Certification: October 1, 2025

Certification Expires: October 1, 2026



NGSP Steering Committee Chair



NGSP Network Coordinator



SRL director/ supervisor