



Yumizen G PT Reco 10 for Yumizen G Line



Reference No.:		1300036376		
Lot No.:		950919 (rev. 1)		
Expiry date:		2027-09-30		
Coagulometers		ISI	MASTER CURVE	
Optical	Yumizen G Line		%	SEC
		100	10,4	
		50	14,3	
		33	18,4	
		25	22,2	
		1,07	MNPT	10,4
Important notice				
<p>The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.</p> <p>For Yumizen G200, Yumizen G400 and Yumizen G400 DDi, the Min. Time must be set to 7 seconds and the Lag time to 5 seconds in the PT setup. Please refer to the Yumizen G PT Reco 10 reagent notice.</p>				