



Yumizen G CTRL I (level 1) for Yumizen G Line



D 4 9 5 0 9 2 4 4 6 4 1 2 0 9 2 7 D



D 5 1 1 0 1 4 9 0 6 8 1 1 3 0 8 5 D



D 6 1 2 7 1 0 6 1 4 3 0 6 9 1 1 5 D



D 7 0 8 4 1 2 5 0 8 6 1 1 7 0 8 2 D



D 8 1 3 7 0 7 7 1 1 6 2 9 0 3 9 3 D



D 9 2 7 9 3 7 8 1 9 1 2 8 7 2 1 1 D



D 0 3 1 6 3 8 8 D

Reference No.:		1300036412	
Lot No:		950924 (rev. 2)	
Expiry date:		2027-09-30	
Optical	Yumizen G Line		
Kit	Parameter	Unit	Range
PT	Prothrombin Time	sec	11.0-14.9
	Prothrombin %	%	68-113
	Prothrombin INR	INR	0.85-1.27
PT Liq	Prothrombin Time	sec	10.6-14.3
	Prothrombin %	%	69-115
	Prothrombin INR	INR	0.84-1.25
PT R	Prothrombin Time	sec	8.6-11.7
	Prothrombin %	%	82-137
	Prothrombin INR	INR	0.77-1.16
APTT	Partial Thromboplastin Time	sec	29.0-39.3
APTT Liq	Partial Thromboplastin Time	sec	27.9-37.8
FIB Cl	Fibrinogen	g/L	1.91-2.87
TT	Thrombin Time	sec	21.1-31.6
AT	Antithrombin	%	80-109
F II	Factor II Deficient Plasma	%	N/A-N/A
F V	Factor V Deficient Plasma	%	N/A-N/A
F VII	Factor VII Deficient Plasma	%	N/A-N/A
F VIII	Factor VIII Deficient Plasma	%	N/A-N/A
F IX	Factor IX Deficient Plasma	%	N/A-N/A
F X	Factor X Deficient Plasma	%	N/A-N/A
F XI	Factor XI Deficient Plasma	%	N/A-N/A
F XII	Factor XII Deficient Plasma	%	N/A-N/A
Important notice			
<p>The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.</p> <p>The % and INR dimension are linked to this MNPT. Then, the ranges of the controls need to be adjusted.</p>			



Yumizen G CTRL II (level 2) for Yumizen G Line



D 4 9 5 0 9 2 4 5 6 4 1 2 0 9 2 7 D



D 5 1 8 0 2 7 0 0 2 7 0 4 5 1 6 2 D



D 6 2 4 2 1 7 5 2 6 3 0 2 5 0 4 2 D



D 7 1 7 0 2 5 5 1 4 3 2 1 4 0 2 6 D



D 8 0 4 4 1 3 8 2 0 7 7 0 0 9 4 7 D



D 9 6 7 4 9 1 2 0 6 7 1 5 5 0 0 0 D



D 0 9 9 9 3 7 3 D

Reference No.:		1300036412	
Lot No:		950924 (rev. 2)	
Expiry date:		2027-09-30	
Optical	Yumizen G Line		
Kit	Parameter	Unit	Range
PT	Prothrombin Time	sec	18.0-27.0
	Prothrombin %	%	27-45
	Prothrombin INR	INR	1.62-2.42
PT Liq	Prothrombin Time	sec	17.5-26.3
	Prothrombin %	%	25-42
	Prothrombin INR	INR	1.70-2.55
PT R	Prothrombin Time	sec	14.3-21.4
	Prothrombin %	%	26-44
	Prothrombin INR	INR	1.38-2.07
APTT	Partial Thromboplastin Time	sec	70.0-94.7
APTT Liq	Partial Thromboplastin Time	sec	67.4-91.2
FIB Cl	Fibrinogen	g/L	0.67-1.55
TT	Thrombin Time	sec	N/A
AT	Antithrombin	%	33-56
F II	Factor II Deficient Plasma	%	N/A-N/A
F V	Factor V Deficient Plasma	%	N/A-N/A
F VII	Factor VII Deficient Plasma	%	N/A-N/A
F VIII	Factor VIII Deficient Plasma	%	N/A-N/A
F IX	Factor IX Deficient Plasma	%	N/A-N/A
F X	Factor X Deficient Plasma	%	N/A-N/A
F XI	Factor XI Deficient Plasma	%	N/A-N/A
F XII	Factor XII Deficient Plasma	%	N/A-N/A
Important notice			
<p>The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.</p> <p>The % and INR dimension are linked to this MNPT. Then, the ranges of the controls need to be adjusted.</p>			