







# ABX Minilyse LMG 1L

## Внутренний контроль качества

Контрольные образцы крови HORIBA необходимо использовать для периодической оценки качества функционирования реагентов и аппарата в заданных диапазонах.

HORIBA предлагает Онлайн-программу межлабораторного сравнения (QCP), которая предоставляет доступ по Интернету к:

- подача отчетов о результатах внутреннего контроля качества в онлайн-режиме;
- мониторинг качества работы аналитического оборудования и непосредственное сравнение результатов исследований с результатами, полученными в сотнях лабораторий по всему миру;
- получение в режиме реального времени статистических отчетов групп коллег через QCP.

Более подробную информацию можно получить на веб-сайте по адресу:

<http://qcp.horiba-abx.com>

## Прослеживаемость калибраторов и контрольных материалов

Не применимо.

## Референсные интервалы

Не применимо.

## Литература

1. US Department of labor, Occupational Safety and Health Administration. 29 CFR 1910. 1030: Occupational Safety and Health Standards: Bloodborne pathogens.
2. Council Directive (2000/54/EC). Official Journal of the European Communities. No. L262 from October 17, 2000: 21-45.
3. Protection of Laboratory Workers From Occupationally Acquired Infections; Approved Guideline - Fourth Edition. CLSI (NCCLS), document M29-A4 (2014) **34** (18).
4. Collection of Diagnostic Venous Blood Specimens - Seventh Edition. CLSI (NCCLS), document GP41 (2017).