

# Yumizen G PT 5 for Yumizen G Line



<b>Reference No.:</b>		<b>1300036338</b>		
<b>Lot No.:</b>		<b>951109 (rev. 1)</b>		
<b>Expiry date:</b>		<b>2027-11-30</b>		
<b>Coagulometers</b>		<b>ISI</b>	<b>MASTER CURVE</b>	
<b>Optical</b>	<b>Yumizen G Line</b>		<b>1,24</b>	<b>%</b>
		<b>100</b>		<b>12,4</b>
		<b>50</b>		<b>17,3</b>
		<b>33</b>		<b>22,4</b>
		<b>25</b>		<b>27,2</b>
			<b>MNPT</b>	<b>12,4</b>
<b>Important notice</b>				
<ul style="list-style-type: none"> <li>- The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.</li> <li>- The efficiency of the reconstitution of the reagent may be improved by performing it at 37°C.</li> </ul>				