



HORIBA ABX SAS
 Parc Euromédecine
 Rue du Caducée
 B.P. 7290
 34184 Montpellier cedex 4
 FRANCE
 www.horiba-abx.com



A03A00009B

Yumizen G PT Liq 4 for Yumizen G Line



Reference No.:		1300036373		
Lot No.:		951111 (rev. 1)		
Expiry date:		2027-05-31		
Coagulometers		ISI	MASTER CURVE	
Optical	Yumizen G Line		1,25	%
		100		11,9
		50		17,5
		33		23,4
		25		28,8
			MNPT	11,9
Important notice				
The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.				