



HORIBA ABX SAS
 Parc Euromédecine
 Rue du Caducée
 B.P. 7290
 34184 Montpellier cedex 4
 FRANCE
 www.horiba-abx.com



A03A00012B

Yumizen G APTT Liq 4 for YUMIZEN G Line



D 4 9 5 1 2 0 7 3 6 3 8 1 0 6 2 7 D



D 5 2 6 6 0 8 3 D

Reference No.:		1300036381
Lot No.:		951207 (rev. 1)
Expiry date:		2027-06-30
MNAPTT		
Coagulometers		Sec
Optical	Yumizen G Line	26,6
Important notice		
<p>The Mean Normal Activated Partial Thromboplastin Time (MNAPTT) is for information only. The value depends on the population, race, gender, sampling tube, etc. According to the CLSI every laboratory should determine its local MNAPTT.</p>		