



Multi-Parameter Assayed Hematology Calibrator

Rx Only

Open-vial stability 5 days

Lot No.: 60611600
Exp. Date: 2026-04-15

Instrument:
Performed by:
Date:

Table of Assigned Values	WBC K/ μ L	RBC M/ μ L	HGB g/dL	HCT %	PLT K/ μ L	MPV fL
ABX MICROS 60	8.4 \pm 0.3	4.21 \pm 0.10	11.3 \pm 0.2	33.7 \pm 1.0	204 \pm 15	9.0 \pm NA
ABX MICROS ES 60	8.3 \pm 0.3	4.18 \pm 0.10	11.3 \pm 0.2	33.2 \pm 1.0	202 \pm 15	9.0 \pm NA
ABX MICROS ESV 60	8.1 \pm 0.5	4.17 \pm 0.24	11.0 \pm 0.3	34.3 \pm 1.9	203 \pm 36	10.5 \pm NA

Table 1. Determine Mean Value	WBC K/ μ L	RBC M/ μ L	HGB g/dL	HCT %	PLT K/ μ L
Sample 1					
2					
3					
4					
5					
6					
Total					
Mean Value (Total 6)					
SD					
%CV (SD Mean Value x 100)					

Table 2. Determine Calibration Factor	WBC K/ μ L	RBC M/ μ L	HGB g/dL	HCT %	PLT K/ μ L
<u>ABX Minocal Assigned Value</u> Mean Value	_____	_____	_____	_____	_____
Calibration Factor (round off to 3 decimal places)					
ABX Minocal Reading (Table 1, Sample 6)					
Corrected Calibrator Reading for Adjustment					

Table 3. Verify Calibration	WBC K/ μ L	RBC M/ μ L	HGB g/dL	HCT %	PLT K/ μ L
Sample 1					
2					
3					
Total					
Mean Value (Total 3)					

For technical assistance, call the Customer Support Center at 1-888-903-5001.