

Yumizen G CTRL I (level 1) for Yumizen G Line



D 4 9 5 1 2 2 4 4 6 4 1 2 1 2 2 7 D



D 5 1 0 8 1 4 6 0 7 2 1 1 9 0 8 2 D



D 6 1 2 3 1 0 1 1 3 7 0 7 2 1 2 1 D



D 7 0 8 2 1 2 2 0 8 7 1 1 8 0 7 3 D



D 8 1 2 1 0 8 1 1 2 2 3 2 6 4 4 1 D



D 9 2 5 4 3 4 3 1 9 8 2 9 7 1 9 2 D



D 0 2 8 8 3 4 8 D

Reference No.:		1300036412	
Lot No:		951224 (rev. 2)	
Expiry date:		2027-12-31	
Optical	Yumizen G Line		
Kit	Parameter	Unit	Range
PT	Prothrombin Time	sec	10.8-14.6
	Prothrombin %	%	72-119
	Prothrombin INR	INR	0.82-1.23
PT Liq	Prothrombin Time	sec	10.1-13.7
	Prothrombin %	%	72-121
	Prothrombin INR	INR	0.82-1.22
PT R	Prothrombin Time	sec	8.7-11.8
	Prothrombin %	%	73-121
	Prothrombin INR	INR	0.81-1.22
APTT	Partial Thromboplastin Time	sec	32.6-44.1
APTT Liq	Partial Thromboplastin Time	sec	25.4-34.3
FIB Cl	Fibrinogen	g/L	1.98-2.97
TT	Thrombin Time	sec	19.2-28.8
AT	Antithrombin	%	81-110
F II	Factor II Deficient Plasma	%	N/A
F V	Factor V Deficient Plasma	%	N/A
F VII	Factor VII Deficient Plasma	%	N/A
F VIII	Factor VIII Deficient Plasma	%	N/A
F IX	Factor IX Deficient Plasma	%	N/A
F X	Factor X Deficient Plasma	%	N/A
F XI	Factor XI Deficient Plasma	%	N/A
F XII	Factor XII Deficient Plasma	%	N/A
Important notice			
<p>The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.</p> <p>The % and INR dimension are linked to this MNPT. Then, the ranges of the controls need to be adjusted.</p>			



Yumizen G CTRL II (level 2) for Yumizen G Line



D 4 9 5 1 2 2 4 5 6 4 1 2 1 2 2 7 D



D 5 1 7 8 2 6 7 0 2 7 0 4 5 1 5 9 D



D 6 2 3 9 1 6 5 2 4 8 0 2 8 0 4 7 D



D 7 1 5 8 2 3 7 1 4 4 2 1 6 0 2 6 D



D 8 0 4 4 1 4 0 2 1 1 6 0 0 8 1 1 D



D 9 6 1 3 8 2 9 0 7 4 1 7 2 0 0 0 D



D 0 9 9 9 3 7 9 D

Reference No.:		1300036412	
Lot No:		951224 (rev. 2)	
Expiry date:		2027-12-31	
Optical	Yumizen G Line		
Kit	Parameter	Unit	Range
PT	Prothrombin Time	sec	17.8-26.7
	Prothrombin %	%	27-45
	Prothrombin INR	INR	1.59-2.39
PT Liq	Prothrombin Time	sec	16.5-24.8
	Prothrombin %	%	28-47
	Prothrombin INR	INR	1.58-2.37
PT R	Prothrombin Time	sec	14.4-21.6
	Prothrombin %	%	26-44
	Prothrombin INR	INR	1.40-2.11
APTT	Partial Thromboplastin Time	sec	60.0-81.1
APTT Liq	Partial Thromboplastin Time	sec	61.3-82.9
FIB Cl	Fibrinogen	g/L	0.74-1.72
TT	Thrombin Time	sec	N/A
AT	Antithrombin	%	37-62
F II	Factor II Deficient Plasma	%	N/A
F V	Factor V Deficient Plasma	%	N/A
F VII	Factor VII Deficient Plasma	%	N/A
F VIII	Factor VIII Deficient Plasma	%	N/A
F IX	Factor IX Deficient Plasma	%	N/A
F X	Factor X Deficient Plasma	%	N/A
F XI	Factor XI Deficient Plasma	%	N/A
F XII	Factor XII Deficient Plasma	%	N/A
Important notice			
<p>The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.</p> <p>The % and INR dimension are linked to this MNPT. Then, the ranges of the controls need to be adjusted.</p>			