

# Yumizen G DDi 2 for Yumizen G15X0x and Yumizen G405

<b>Reference No.:</b>		<b>1300036391</b>	
<b>Lot No.:</b>		<b>951124 (rev. 1)</b>	
<b>Expiry date:</b>		<b>2026-11-30</b>	
<b>Coagulometers</b>		<b>MASTER CURVE</b>	
<b>Optical</b>	<b>Yumizen G15X0x/G405</b>	<b>µg(FEU)/mL</b>	<b>ΔOD</b>
		4.00	0.2230
		3.00	0.1690
		2.00	0.1140
		1.00	0.0585
		0.50	0.0310
		0.25	0.0175
<b>Slope value</b>		<b>0,07</b>	



# Yumizen G DDi 2 for Yumizen G8X0x

<b>Reference No.:</b>		<b>1300036391</b>	
<b>Lot No.:</b>		<b>951124 (rev. 1)</b>	
<b>Expiry date:</b>		<b>2026-11-30</b>	
<b>Coagulometers</b>		<b>MASTER CURVE</b>	
<b>Optical</b>	<b>Yumizen G8X0x</b>	<b>µg(FEU)/mL</b>	<b>ΔOD</b>
		4.00	0.2333
		3.00	0.1792
		2.00	0.1161
		1.00	0.0551
		0.50	0.0285
		0.25	0.0162
<b>Slope value</b>		<b>0,06</b>	





# Yumizen G DDi 2 for Yumizen G400 DDi/G200

<b>Reference No.:</b>		<b>1300036391</b>	
<b>Lot No.:</b>		<b>951124 (rev. 1)</b>	
<b>Expiry date:</b>		<b>2026-11-30</b>	
<b>Coagulometers</b>		<b>MASTER CURVE</b>	
		<b>µg(FEU)/mL</b>	<b>ΔOD</b>
<b>Optical</b>	<b>Yumizen G400 DDi/G200</b>	<b>4.00</b>	<b>0.0855</b>
		<b>3.00</b>	<b>0.0665</b>
		<b>2.00</b>	<b>0.0455</b>
		<b>1.00</b>	<b>0.0240</b>
		<b>0.50</b>	<b>0.0135</b>
		<b>0.25</b>	<b>0.0085</b>
		<b>Slope value</b>	<b>0,03</b>
<b>Important notice</b>			
<p>For Yumizen G400 DDi and Yumizen G200, make sure that the end time is set to 180 seconds in the D-Dimer setup and the curve type <b>Lin-Lin p-p</b> is selected in <b>Calibration &gt; D-Dimer &gt; ug FEU/ml</b>.          For more information, please contact your local HORIBA Medical representative.</p>			