



**HORIBA ABX SAS**  
 Parc Euromédecine  
 Rue du Caducée  
 B.P. 7290  
 34184 Montpellier cedex 4  
 FRANCE  
 www.horiba-abx.com



A03A00009B

# Yumizen G PT Liq 4 for Yumizen G Line



D 4 9 5 1 2 1 7 3 6 3 7 3 0 6 2 7 D



D 5 1 0 0 1 1 6 5 0 0 1 7 1 3 3 0 D



D 6 2 2 8 2 5 0 2 8 1 1 1 6 1 2 3 D



D 7 1 5 7 D

<b>Reference No.:</b>		<b>1300036373</b>		
<b>Lot No.:</b>		<b>951217 (rev. 1)</b>		
<b>Expiry date:</b>		<b>2027-06-30</b>		
<b>Coagulometers</b>		<b>ISI</b>	<b>MASTER CURVE</b>	
<b>Optical</b>	<b>Yumizen G Line</b>		<b>1,23</b>	<b>%</b>
		<b>100</b>		<b>11,6</b>
		<b>50</b>		<b>17,1</b>
		<b>33</b>		<b>22,8</b>
		<b>25</b>		<b>28,1</b>
	<b>MNPT</b>	<b>11,6</b>		
<b>Important notice</b>				
The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.				