



Yumizen G CAL

for Yumizen G Line

| Reference No.: | | 1300036416 |
|--|-------------|------------------------|
| Lot No.: | | 950520 (rev. 1) |
| Expiry date: | | 2027-05-31 |
| Parameter | Unit | Value |
| PT | % | 83 |
| PT Liq | % | 83 |
| PT R | % | 84 |
| Fibrinogen | g/L | 3.01 |
| | mg/dL | 301 |
| Antithrombin | % | 109 |
| | IU/mL | 1.09 |
| Factor II Deficient Plasma | % | 85 |
| | IU/mL | 0.85 |
| Factor V Deficient Plasma | % | 75 |
| | IU/mL | 0.75 |
| Factor VII Deficient Plasma | % | 95 |
| | IU/mL | 0.95 |
| Factor VIII Deficient Plasma | % | 80 |
| | IU/mL | 0.80 |
| Factor IX Deficient Plasma | % | 97 |
| | IU/mL | 0.97 |
| Factor X Deficient Plasma | % | 90 |
| | IU/mL | 0.90 |
| Factor XI Deficient Plasma | % | 84 |
| | IU/mL | 0.84 |
| Factor XII Deficient Plasma | % | 87 |
| | IU/mL | 0.87 |
| Important notice | | |
| Dilution serie must be prepared. Please refer to the notice. | | |