



# *Yumizen G TT for Yumizen G Line*



<b>Reference No.:</b>		<b>1300036382</b>
<b>Lot No.:</b>		<b>960216</b>
<b>Expiry date:</b>		<b>2028-02-29</b>
<b>MNTT</b>		
<b>Coagulometers</b>		<b>sec</b>
<b>Optical</b>	<b>Yumizen G Line</b>	<b>17,6</b>
<b>Important notice</b>		
<p>The Mean Normal Thrombin Time (MNTT) is for information only. The value depends on the population, race, gender, sampling tube etc. According to the CLSI every laboratory should determine its local MNTT.</p>		