

LOT CX 519
 Rev 1

CAL

(Exp.) 2026-07-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	
		EMI CRP	LC-767G	MS CRP	XLR	DX NEXUS DF NEXUS	MS60		60 60 C+ ES60	XL80	
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	9.2	9.0	9.2	9.1	9.4	9.1		9.2	9.1	± 0.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	4.34	4.33	4.30	4.36	4.39	4.34		4.39	4.36	± 0.06
HB HGB	g/dl	12.6	12.6	12.7	12.8	12.8	12.8		12.9	12.8	± 0.2
	g/l	126	126	127	128	128	128		129	128	± 2
	mmol/l	7.82	7.82	7.89	7.95	7.95	7.95		8.01	7.95	± 0.12
HT HCT	%	36.0	35.5	34.4	35.1	34.9	34.5		35.3	35.1	± 1.0
	l/l	0.360	0.355	0.344	0.351	0.349	0.345		0.353	0.351	± 0.010
PLA PLT	10 ³ /mm ³ ;10 ⁹ /l	232	237	232	230	241	231		230	230	± 10
VMP MPV	µm ³ ; fl	8.3	8.2	N/A	N/A	N/A	N/A		N/A	N/A	± 0.5

Ref: TEMP-0387 Rev.53