

LOT

CX 520

CAL



(Exp.) 2026-08-05
 (YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	
		EMI CRP	LC-767G	MS CRP	XLR	DX NEXUS DF NEXUS	MS60		60 60 C+ ES60	XL80	
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	8.9	8.9	8.9	8.7	9.1	8.7		9.0	8.7	± 0.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	4.50	4.51	4.50	4.50	4.56	4.52		4.55	4.50	± 0.06
HB HGB	g/dl	13.0	13.1	13.2	13.2	13.1	13.1		13.3	13.2	± 0.2
	g/l	130	131	132	132	131	131		133	132	± 2
	mmol/l	8.07	8.14	8.20	8.20	8.14	8.14		8.26	8.20	± 0.12
HT HCT	%	37.1	37.0	35.8	36.2	36.0	35.7		36.6	36.2	± 1.0
	l/l	0.371	0.370	0.358	0.362	0.360	0.357		0.366	0.362	± 0.010
PLA PLT	10 ³ /mm ³ ;10 ⁹ /l	227	225	224	224	228	228		228	224	± 10
VMP MPV	µm ³ ; fl	8.1	8.0	N/A	N/A	N/A	N/A		N/A	N/A	± 0.5

Ref: TEMP-0387 Rev.53