

Yumizen G CTRL I (level 1) for Yumizen G Line



Reference No.:		1300036412	
Lot No:		960330 (rev. 2)	
Expiry date:		2028-03-31	
Optical	Yumizen G Line		
Kit	Parameter	Unit	Range
PT	Prothrombin Time	sec	9.7-13.1
	Prothrombin %	%	100-166
	Prothrombin INR	INR	0.71-1.06
PT Liq	Prothrombin Time	sec	9.4-12.7
	Prothrombin %	%	95-158
	Prothrombin INR	INR	0.72-1.08
PT R	Prothrombin Time	sec	7.7-10.4
	Prothrombin %	%	103-172
	Prothrombin INR	INR	0.70-1.06
APTT	Partial Thromboplastin Time	sec	33.7-45.7
APTT Liq	Partial Thromboplastin Time	sec	27.2-36.8
FIB Cl	Fibrinogen	g/L	2.21-3.31
TT	Thrombin Time	sec	19.6-29.5
AT	Antithrombin	%	82-111
F II	Factor II Deficient Plasma	%	N/A
F V	Factor V Deficient Plasma	%	N/A
F VII	Factor VII Deficient Plasma	%	N/A
F VIII	Factor VIII Deficient Plasma	%	N/A
F IX	Factor IX Deficient Plasma	%	N/A
F X	Factor X Deficient Plasma	%	N/A
F XI	Factor XI Deficient Plasma	%	N/A
F XII	Factor XII Deficient Plasma	%	N/A
Important notice			
<p>The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.</p> <p>The % and INR dimension are linked to this MNPT. Then, the ranges of the controls need to be adjusted.</p>			

Yumizen G CTRL II (level 2) for Yumizen G Line



Reference No.:		1300036412	
Lot No:		960330 (rev. 2)	
Expiry date:		2028-03-31	
Optical	Yumizen G Line		
Kit	Parameter	Unit	Range
PT	Prothrombin Time	sec	18.6-27.9
	Prothrombin %	%	26-43
	Prothrombin INR	INR	1.68-2.52
PT Liq	Prothrombin Time	sec	16.4-24.6
	Prothrombin %	%	28-46
	Prothrombin INR	INR	1.63-2.44
PT R	Prothrombin Time	sec	14.6-22.0
	Prothrombin %	%	25-42
	Prothrombin INR	INR	1.43-2.14
APTT	Partial Thromboplastin Time	sec	67.4-91.2
APTT Liq	Partial Thromboplastin Time	sec	64.7-87.5
FIB Cl	Fibrinogen	g/L	0.74-1.73
TT	Thrombin Time	sec	N/A
AT	Antithrombin	%	36-60
F II	Factor II Deficient Plasma	%	N/A
F V	Factor V Deficient Plasma	%	N/A
F VII	Factor VII Deficient Plasma	%	N/A
F VIII	Factor VIII Deficient Plasma	%	N/A
F IX	Factor IX Deficient Plasma	%	N/A
F X	Factor X Deficient Plasma	%	N/A
F XI	Factor XI Deficient Plasma	%	N/A
F XII	Factor XII Deficient Plasma	%	N/A
Important notice			
<p>The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.</p> <p>The % and INR dimension are linked to this MNPT. Then, the ranges of the controls need to be adjusted.</p>			