



HORIBA ABX SAS

BP 7290

34184

Montpellier

cedex 4 -

France

www.horiba-abx.com/documentation



A03A00001A

Yumizen G APTT 4 for Yumizen G Line



D4960416363770428D



D5334085D

Reference No.:		1300036377
Lot No.:		960416 (rev.1)
Expiry date:		2028-04-30
MNAPTT		
Coagulometers		Sec
Optical	Yumizen G Line	33,4
Important notice		
The Mean Normal Activated Partial Thromboplastin Time (MNAPTT) is for information only. The value depends on the population, race, gender, sampling tube, etc. According to the CLSI every laboratory should determine its local MNAPTT.		