

REF 1300159585

REAGENT 1 L

 **HORIBA ABX SAS**
Parc Euromédecine
Rue du Caducée
BP 7290
34184 Montpellier Cedex 4
FRANCE

Yumivet ProClean 1L

- Yumivet VH2500

Hematology Devices (for *in vitro* diagnostic use)

Intended Use

The **Yumivet ProClean 1L** is an enzymatic solution with proteolytic action intended for *in vitro* diagnostic use and designed for the cleaning of HORIBA blood cell counters. Veterinary use only.

Warnings and Precautions

- **Yumivet ProClean 1L** is for professional *in vitro* diagnostic use only.
For veterinary laboratory use.
- It is the user's responsibility to verify that this document is applicable to the product use.
- This reagent is classified as hazardous in compliance with regulation (EC) N°.1272/2008.
- **Warning**
H317: May cause an allergic skin reaction.
H412: Harmful to aquatic life with long lasting effects.
P261: Avoid breathing dust/fume/gas/mist/vapours/spray.
P280: Wear protective gloves or clothing and eye or face protection.
P302 + P352: IF ON SKIN: Wash with plenty of soap and water.
P333 + P313: If skin irritation or rash occurs: Get medical advice/attention.
P501: Dispose of contents and container in accordance with all local, regional, national and international regulations.
Contains: 2-methyl-(2H)-isothiazol-3-one
Contains: Octhilinone (ISO)
- Users are advised to wear approved protective clothing when handling chemical products: lab coat, gloves, and eye protection.
- Observe the standard laboratory precautions for use and follow national or local health and safety guidelines.
- User must be trained by a HORIBA representative before attempting to operate the device.

- In the event of a malaise following skin contact, ingestion, or inhalation, consult a doctor.
- Please refer to the Safety Data Sheet (SDS) associated with **Yumivet ProClean 1L**.
- Do not use the product if the recommended storage conditions, including temperature, are not followed.
- Any serious incident that has occurred in relation to the device shall be reported to the manufacturer and the competent authority of the country in which the user and/or the patient is established.
- Harmful to the environment.
Avoid release to the environment.
- The reagent containers are disposable and should be disposed of in accordance with the local legal requirements.
- For technical assistance, you can call +33 (0)4 67 14 15 16.
- This reagent is destined for use with HORIBA blood cell counters specified above. HORIBA cannot guarantee the correct functioning of this reagent with instruments other than those specified above, or with instruments not manufactured by HORIBA.

Waste Management

Please refer to local legal requirements.
This reagent contains less than 0.1% of sodium azide as a preservative.

Microbiological State

Not applicable.

Yumivet ProClean 1L

Description and Composition

Description:

Limpid and colourless to light yellowish aqueous solution.

Composition:

Organic buffer	< 5%
Proteolytic enzyme	< 1%
Preservative	< 1%

Storage and Stability ^a

- **Storage condition (before opening):** 18-25°C (65-77°F).
Do not freeze.
- **Open stability:** 3 months maximum at 15-30°C (59-86°F) after opening and within the expiration limit.
- **Expiration date:** refer to "expiration date" reagent packaging label.

Materials Required but not Provided

- Veterinary automated hematology analyzer.
- Standard laboratory equipment.

Specimen

Not applicable.

Procedure

This reagent is ready to use.

Procedure for plugged-in reagent

Plugged-in reagent is used on:

- Yumivet VH2500
1. If necessary, remove the empty **Yumivet ProClean 1L** from the reagent compartment.
 2. Refer to the user manual to identify **Yumivet ProClean 1L** using the barcode reader or manually.

3. Install **Yumivet ProClean 1L** into the reagent compartment of the instrument.
4. Gently push it down in order to plug it correctly into the male connectors.

Follow instructions displayed on your instrument software.
Refer to the instrument user manual for detailed analysis and control procedures.

Methodology

Yumivet ProClean 1L: the combined action of a proteolytic enzyme with a detergent eliminates protein residues and prevents the hydraulic tubes from clogging and / or blocking. It is used also to break down the protein build-ups in the counting chambers and apertures.

Performance Characteristics and Limitations

Not applicable.

Calculation and Interpretation of Analytical Results

Not applicable.

Changes in the Procedure and in the Performance

Packaging spoiling

In case of protective packaging spoiling, do not use **Yumivet ProClean 1L** if the damages might have an effect on the product performance.

Signs of deterioration

In the event of any signs of physical or chemical deterioration (turbidity, change in colour etc.) **Yumivet ProClean 1L** should be replaced.

Temperature limits

Do not use the **Yumivet ProClean 1L** if it has been stored at temperatures outside the recommended storage conditions.

^aModification: modification of Storage and Stability.

Yumivet ProClean 1L

Before using **Yumivet ProClean 1L**, make sure it has reached the operating temperature conditions as described in the instrument user manual.

Internal Quality Control

Not applicable.

Traceability of Calibrators and Control Materials

Not applicable.

Reference Intervals

Not applicable.

