

Hematology Devices (for in vitro diagnostic use)

ABX Diluent

23/03/07
A95A00005DEN


Exclusive use:

Micros
Micros CRP/CRP 200
Pentra 60
Pentra 60 C+
Pentra 80
Pentra XL 80
Pentra 120
Pentra 120 Retic
Pentra DF 120 (20L only)
Pentra DX 120 (20L only)
Slide Preparation System

REF 0901020 (20L)
0901010 (10L)*

REAGENT 20L
10L*

IVD 

 **HORIBA ABX**
BP 7290
34184 Montpellier
cedex 4 - France

* 10L ABX Diluent exclusive use

	Software version
Pentra 60	≥ v2.2.0
Pentra 60C+	≥ v2.4.0
Pentra 80	≥ v1.7.1
Pentra XL 80	≥ v1.6.1
Pentra 120	≥ v4.8.0
Pentra 120 Retic	≥ v4.8.0

1. Functions

Buffered isotonic solution for sheating and diluting leucocytes, and for the determination and differentiation of blood cells, and the measurement of hematocrit on HORIBA ABX blood cell counters ^a.

Measurement procedure to be followed in using the device:

Principle of the method, specific analytical performance characteristics, analytical sensitivity, diagnostic sensitivity, analytical specificity, diagnostic specificity, accuracy, repeatability, reproducibility (including control of known relevant interference), limits of detection, limitations of the method and information about the use of available reference measurement procedures and materials by the user: see «Section: Specifications» in the instrument User Manual.

2. Conservation & expiration

Storage conditions: Stored at 18 to 25°C and away from the light.

Expiration date: refer to «expiration date» reagent packaging label.

3. Measurements, principles & results

Directions for use: see «Section: Maintenance & Troubleshooting» in the instrument User Manual.

Measuring Principles : see «Section: Description & technology» in the instrument User Manual.

Results: see «Section: Workflow» in the instrument User Manual

Performance data: see «Section: Specifications» in the instrument User Manual.

Note: if performance changes, call your HORIBA ABX representative.

4. Composition & handling precautions

Composition:

Sodium Chloride.... < 1 %
Natriumazid < 0,1 %
Surfactant. < 0,1 %

pH: 8,1 +/- 0,2 (T = 20°C)

Resistivity: 60,5 +/- 1 Ω (T = 20°C)

Description: Limpid and odorless aqueous solution.

Handling Precautions: Avoid contact with eyes, skin and clothing. Wear laboratory gloves when handling the product. Keep the bottle

a.Modifications from C to D: ABX Diluent 10L

closed when not in use. Please refer to the MSDS associated with the reagent.

Specimen Collection and Mixing: see «Section: Workflow» in the instrument User Manual.

5. Limitations & waste disposal

Limitations: see «Section: Specifications» in the instrument User Manual

Safe Waste Disposal: see «Section: Specifications» in the instrument User Manual. Please refer to the MSDS associated with the reagent.