

# Hematology ABX Devices (for in vitro diagnostic use)

## ABX Minidil UMG

### Exclusive use:

Hematology ABX device	Exclusive
Argos / Helios (3&5DIFF)	
Micros 60 (18P)	✓
Micros CRP	✓
Pentra 60	
Pentra 60 C+	
Pentra 80	
Pentra XL 80	
Pentra 120	
Pentra 120 Retic	
Pentra DX 120	
Slide Preparation System	

15/09/03  
A95A00006-A

**REF** 0802010 (10L)

**REAGENT** 10L

**IVD** 



**ABX Diagnostics**  
BP 7290 - 34187 Montpellier  
cedex 4 - France

## 1. Functions

Buffered isotonic solution for the determination of blood cell counting and the measurement of hematocrit on ABX blood cell counters

### Measurement procedure to be followed in using the device:

Principle of the method, specific analytical performance characteristics, analytical sensitivity, diagnostic sensitivity, analytical specificity, diagnostic specificity, accuracy, repeatability, reproducibility (including control of known relevant interference), limits of detection, limitations of the method and information about the use of available reference measurement procedures and materials by the user: see «Section: Specifications» in the instrument User Manual.

## 2. Conservation & expiration

**Storage conditions:** Stored at 18 to 25°C and away from the light.

**Expiration date:** refer to «expiration date» reagent packaging label.

## 3. Measurements, principles & results

**Directions for use:** see «Reagent Location and connection» in the instrument User Manual.

**Measuring Principles :** see «Measuring Principles» in the instrument User Manual.

**Results:** Refer to the instrument User Manual

**Performance data:** see «Section: Specifications» in the instrument User Manual.

Note: if performance changes, call your ABX Diagnostics representative.

## 4. Composition & handling precautions

### Composition:

Sodium fluoride . . . . . < 3%  
Sodium Azide. . . . . < 0,1%  
Sodium Hydroxyde. . . . . < 1%  
Dimethylourea . . . . . 0,1%

**pH:** 7.0 +/- 0.1 (T = 20°C)

**Resistivity:** 60 +/- 6 Ω (T = 20°C)

**Description:** Limpid and odourless aqueous solution.

**Handling Precautions:** Avoid contact with eyes, skin and clothing. Wear laboratory gloves when handling the product. Keep the bottle closed when not in use. Please refer to the MSDS associated with the reagent.

**Specimen Collection and Mixing:** see «Section: Preparation before analysis and Startup» in the instrument User Manual.

## 5. Limitations & waste disposal

**Limitations:** see «Section: Specifications» in the instrument User Manual

**Safe Waste Disposal:** see «Section: Installation» in the instrument User Manual. Please refer to the MSDS associated with the reagent.

