

Hematology Devices (for in vitro diagnostic use)

ABX Minidil LMG

04/02/09
A95A00006EEN

REF 0802010 (10L)

REAGENT 10L

IVD CE



HORIBA ABX SAS
BP 7290 - 34184 Montpellier
cedex 4 - France

Exclusive use:

ABX Micros 45
ABX Micros 60
ABX Micros ES60
ABX Micros CRP
ABX Micros CRP200
ABX Pentra 60
ABX Pentra 60 C+
ABX Pentra 80
ABX Pentra XL 80
ABX Pentra 120
ABX Pentra 120 Retic
ABX Pentra DX 120

1. Functions

Buffered isotonic solution for the determination of blood cell counting and the measurement of hematocrit on HORIBA Medical blood cell counters^a.

Measurement procedure to be followed in using the device:

Principle of the method, specific analytical performance characteristics, analytical sensitivity, diagnostic sensitivity, analytical specificity, diagnostic specificity, accuracy, repeatability, reproducibility (including control of known relevant interference), limits of detection, limitations of the method and information about the use of available reference measurement procedures and materials by the user: see «Section: Specifications» in the instrument User Manual.

2. Conservation & expiration

Storage conditions: Stored at 18°C (65°F) to 25°C (77°F) and away from the light.

Open stability: maximum 6 months after a vial has been opened.

Expiration date: refer to «expiration date» reagent packaging label.

3. Measurements, principles & results

Directions for use : see «Section: Maintenance & Troubleshooting» in the instrument User Manual. This reagent is for professional *in-vitro* diagnostic use only.

Measuring Principles : see «Measuring Principles» in the instrument User Manual.

Results: Refer to the instrument User Manual

Performance data: see «Section: Specifications» in the instrument User Manual.

Note: if performance changes, call your HORIBA Medical representative.

4. Composition & handling precautions

Composition:

Sodium Fluoride..... < 3%
Sodium Azide..... < 0,1%
Sodium Hydroxyde.< 1%
Dimethylolurea0,1%

pH: 7.0 +/- 0.1 (T = 20°C)

Resistivity: 60 +/- 6 Ω (T = 20°C)

Description: Limpid and odorless aqueous solution.

Handling Precautions: Avoid contact with eyes, skin and clothing. Wear laboratory gloves when handling the product. Keep the bottle closed when not in use. Please refer to the MSDS associated with the reagent.

a.Modifications from index D to E: ABX Micros ES60

Specimen Collection and Mixing: see «Section: Preparation before analysis and Startup» in the instrument User Manual.

5. Limitations & waste disposal

Limitations: see «Section: Specifications» in the instrument User Manual

Safe Waste Disposal: see «Section: Installation» in the instrument User Manual. Please refer to the MSDS associated with the reagent.