

Hematology Devices (for in vitro diagnostic use)

ABX Cleaner



(From lot #091209 and above)

08/12/09
A95A00012IEN

Exclusive use:

ABX Micros ES60*
ABX Micros CRP 200*
ABX Pentra 60
ABX Pentra 60 C+
Pentra ES60
ABX Pentra 80
ABX Pentra XL 80
ABX Pentra 120
ABX Pentra 120 Retic
ABX Pentra DF 120
ABX Pentra DX 120

REF 0903010 (1L)
0903011 (0,5L)*

REAGENT 1L
0,5L*

IVD



HORIBA ABX SAS
BP 7290 - 34184 Montpellier
cedex 4 - France

1. Functions

Enzymatic solution with proteolytic action for the cleaning of HORIBA Medical blood cell counters.

Measurement procedure to be followed in using the device:

Principle of the method, specific analytical performance characteristics, analytical sensitivity, diagnostic sensitivity, analytical specificity, diagnostic specificity, accuracy, repeatability, reproducibility (including control of known relevant interference), limits of detection, limitations of the method and information about the use of available reference measurement procedures and materials by the user: see «Section: Specifications» in the instrument User Manual.

2. Conservation & Expiration

Storage conditions: Stored at 18°C (65°F) to 25°C (77°F).

Open stability: maximum 1 month after a vial has been opened.

Expiration date: refer to «expiration date» reagent packaging label.

3. Measurements, principles & results

Directions for use : see «Section: Maintenance & Troubleshooting» in the instrument User Manual.

Measuring Principles : see «Section: Description & technology» in the instrument User Manual.

Results: see «Section: Workflow» in the instrument User Manual

Performance data: see «Section: Specifications» in the instrument User Manual.

Note: if performance changes, call your HORIBA Medical representative.

4. Composition & Handling precautions

Composition^a:

Organic Buffer..... < 20 %

Proteolytic Enzyme < 1 %

Sodium Azide..... < 0,1%

pH: 9,9 +/- 0,2 (T = 20°C)

Resistivity: 66,3 +/- 2,0 Ω (T = 20°C)

Description: Transparent liquid.

Handling Precautions: The reagent contains Sodium Azide as preservative. Avoid contact with eyes, skin and clothing. Wear laboratory gloves when handling the product. The product may be harmful if ingested or inhaled. Keep the bottle closed when not in use. Please refer to the MSDS associated with the reagent.

Specimen Collection and Mixing: see «Section: Workflow» in the instrument User Manual.

5. Limitations & waste disposal

Limitations: see «Section: Specifications» in the instrument User Manual.

Safe Waste Disposal: see «Section: Specifications» in the instrument User Manual. Please refer to the MSDS associated with the reagent.

a.Modification from index H to I : Sodium azide added

