

Hematology Devices (for in vitro diagnostic use)

## ABX Miniclean

(From lot #091209 and above)

### Exclusive use:

ABX Micros 45  
ABX Micros 60  
ABX Micros ES60  
ABX Micros CRP/CRP200  
ABX Pentra 60  
ABX Pentra 60 C+  
ABX Pentra 80  
ABX Pentra XL 80  
ABX Pentra 120  
ABX Pentra 120 Retic  
ABX Pentra DX 120  
Slide Preparation System

08/12/09  
A95A00031GEN

**REF** 0403010 (1L)

**REAGENT** 1L

**IVD** 



**HORIBA ABX SAS**  
BP 7290 - 34184 Montpellier  
cedex 4 - France

## 1. Functions

Enzymatic solution with proteolytic action for the cleaning of HORIBA Medical blood cell counters.

### Measurement procedure to be followed in using the device:

Principle of the method, specific analytical performance characteristics, analytical sensitivity, diagnostic sensitivity, analytical specificity, diagnostic specificity, accuracy, repeatability, reproducibility (including control of known relevant interference), limits of detection, limitations of the method and information about the use of available reference measurement procedures and materials by the user: see «Section: Specifications» in the instrument User Manual.

## 2. Conservation & Expiration

**Storage conditions:** Stored at 18°C (65°F) to 25°C (77°F).

**Open stability:** maximum 1 month after a vial has been opened .

**Expiration date:** refer to «expiration date» reagent packaging label.

## 3. Measurements, principles & results

**Directions for use:** see «Section: Maintenance & Troubleshooting» in the instrument User Manual. This reagent is for professional *in-vitro* diagnostic use only.

**Measuring Principles :** see «Section: Technology» in the instrument User Manual.

**Results:** Refer to the instrument User Manual

**Performance data:** see «Section: Specifications» in the instrument User Manual.

Note: if performance changes, call your HORIBA Medical representative.

## 4. Composition & Handling precautions

### Composition<sup>a</sup>:

Organic Buffer..... < 20 %

Proteolytic Enzyme ..... < 1 %

Sodium Azide..... <0,1%

**pH:** 9,9 +/- 0,2 (T = 20°C)

**Resistivity:** 66,3 +/- 2,0 Ω (T = 20°C)

**Description:** Transparent liquid.

**Handling Precautions:** The reagent contains Sodium Azide as preservative. Avoid contact with eyes, skin and clothing. Wear laboratory gloves when handling the product. The product may be harmful if ingested or inhaled. Keep the bottle closed when not in use. Please refer to the MSDS associated with the reagent.

**Specimen Collection and Mixing:** see «Section: Workflow» in the instrument User Manual.

## 5. Limitations & waste disposal

**Limitations:** see «Section: Specifications» in the instrument User Manual.

**Safe Waste Disposal:** Follow your laboratory's protocol when neutralizing and disposing of waste. Please refer to the MSDS associated with the reagent.

a.Modification from index F to G : Sodium azide added

