

Hematology ABX Devices (for in vitro diagnostic use)

ABX Basolyse

Exclusive use:

Hematology ABX device	Exclusive
Argos/Helios (5diff)	✓
Micros 60	
Micros CRP	
Pentra 60	
Pentra 60 C+	
Pentra 80	
Pentra XL 80	
Pentra 120	✓
Pentra 120 Retic	✓
Pentra DX 120	✓
Slide Preparation System	

15/07/03
A95A00001-A

REF 0204050 (5L)

REAGENT 5L

IVD 



ABX Diagnostics
BP 7290 - 34187 Montpellier
cedex 4 - France

1. Functions

Solution for the differentiation of leucocytes subpopulations on ABX blood cell counters.

Measurement procedure to be followed in using the device:

Principle of the method, specific analytical performance characteristics, analytical sensitivity, diagnostic sensitivity, analytical specificity, diagnostic specificity, accuracy, repeatability, reproducibility, limits of detection, limitations of the method and information about the use of available reference measurement procedures and materials by the user: see «Section: Specifications» in the instrument User Manual.

2. Conservation & expiration

Storage conditions: Room temperature between 18°C (65°F) to 25°C (77°F).

Expiration date: refer to «expiration date» reagent packaging label.

3. Measurements, principles & results

Directions for use: see «Section: Maintenance & Troubleshooting / Reagent Location and connection» in the instrument User Manual. This reagent is for professional *in-vitro* diagnostic use only.

Measuring Principles : see «Section: Description & technology» in the instrument User Manual.

Results: see «Section: Workflow» in the instrument User Manual

Performance data: see «Section: Specifications» in the instrument User Manual.

Note: if performance changes, call your ABX Diagnostics representative.

4. Composition & handling precautions

Composition:

Hydrochloric Acid . . . <0.001 %
Phtalic Acid 0.09 %
Sodium Dodecyl Sulfate 0.02 %

pH: 2,7 +/- 0,2 (T = 20°C)

Resistivity: 720 +/- 20 Ω (T = 20°C)

Description: Transparent liquid in transparent plastic 5L cubitainer.

Handling Precautions: Avoid contact with eyes, skin and clothing. Wear laboratory gloves when handling the product. The product may be harmful if ingested or inhaled. Keep the bottle closed when not in use. Avoid contact with products which contains cyanide. Please refer to the MSDS associated with the reagent.

Specimen Collection and Mixing: see «Section: Workflow» in the instrument User Manual.

5. Limitations & waste disposal

Limitations: see «Section: Specifications» in the instrument User Manual.

Safe Waste Disposal: see «Section: Specifications» in the instrument User Manual. Please refer to the MSDS associated with the reagent.

