

HORIBA ABX SAS
Parc Euromédecine
Rue du Caducée – BP7290
34184 MONTPELLIER Cedex 4 – France
Web : www.horiba.com
Tél. : 33 (0) 4 67 14 15 16

Déclaration de Conformité *Declaration of Conformity*

(No.dc 10023c)

Selon
According to ISO 17050-1

NOUS LE MANDATAIRE
WE THE AUTHORISED REPRESENTATIVE

Nom <i>Name</i>	HORIBA ABX SAS
Adresse <i>Address</i>	Parc Euromédecine Rue du Caducée – BP7290 34184 MONTPELLIER Cedex 4 - FRANCE


DECLARONS QUE LE PRODUIT
DECLARE THAT THE PRODUCT

Nom du produit <i>Product name</i>	Hematology Analyzer
Modèles <i>Models</i>	Microsemi CRP
Pays d'origine <i>Country of origin</i>	Japan
Fabricant <i>Manufacturer</i>	HORIBA, Ltd. 2 Miyanogigashi, Kisshoin, Minami-ku, Kyoto 601-8510, Japan

EST CONFORME AUX DIRECTIVES ET NORMES
CONFORMS TO DIRECTIVES AND STANDARDS

Directives <i>Directives</i>	98/79/EC – IVD Medical Devices Classification : Outside Annex II and not for self-testing Conformity Assessment Procedure: Annex III 2002/96/EC - WEEE 2004/108/EC – Electromagnetic compatibility 2006/95/EC – Low Voltage
Normes <i>Standards</i>	IEC61010-1 : 2010 IEC61010-2-081 : 2001 + A1 : 2003 IEC61010-2-101 : 2002 EN61326-2-6 : 2005

Montpellier, France
03 juillet 2013
July 03rd, 2013


Christian DUBUC
Directeur
Senior Director

TEMP-4018 Rev.9