

HORIBA ABX SAS
Parc Euromédecine
Rue du Caducée – BP7290
34184 MONTPELLIER Cedex 4 – France
Web : www.horiba.com
Tél. : 33 (0) 4 67 14 15 16

Déclaration de Conformité *Declaration of Conformity*

(No.de10011i)

Selon
According to ISO 17050-1

NOUS LE FABRICANT **WE THE MANUFACTURER**

Nom <i>Name</i>	HORIBA ABX SAS
Adresse <i>Address</i>	Parc Euromédecine Rue du Caducée – BP7290 34184 MONTPELLIER Cedex 4 - FRANCE

DECLARONS QUE LE PRODUIT **DECLARE THAT THE PRODUCT**

Nom du produit <i>Product name</i>	Hematology analyzer
Modèles <i>Models</i>	ABX MICROS CRP 200
Pays d'origine <i>Country of origin</i>	Japan

EST CONFORME AUX DIRECTIVES ET NORMES **CONFORMS TO DIRECTIVES AND STANDARDS**

Directives <i>Directives</i>	98/79/EC – IVD Medical Devices Classification : Outside Annex II and not for self-testing Conformity Assessment Procedure : Annex III 2006/95/EC – Low voltage
Normes <i>Standards</i>	IEC 61010-1 : 2010 IEC 61010-2-081 : 2001 +A1 : 2003 IEC 61010-2-101 : 2002 EN 61326-1 : 2006 UL 61010-1 : 2004 CAN/CSA-C22.2 61010-1 : 2004

Montpellier, France
26 septembre 2013
September 26th, 2013

Christian DUBUC
Directeur
Senior Director

TEMP-4018 Rev.9

