

HORIBA ABX SAS
Parc Euromédecine
Rue du Caducée – BP7290
34184 MONTPELLIER Cedex 4 – France
Web : www.horiba.com
Tél. : 33 (0) 4 67 14 15 16

Déclaration de Conformité *Declaration of Conformity*

(No.dc40156b)

Selon
According to ISO 17050-1

NOUS LE FABRICANT
WE THE MANUFACTURER

Nom <i>Name</i>	HORIBA ABX SAS
Adresse <i>Address</i>	Parc Euromédecine Rue du Caducée – BP7290 34184 MONTPELLIER Cedex 4 - FRANCE

DECLARONS QUE LE PRODUIT
DECLARE THAT THE PRODUCT

Nom du produit <i>Product name</i>	Self-testing Hematology analyzer
Modèles <i>Models</i>	Micros Care ST

EST CONFORME AUX DIRECTIVES ET NORMES
CONFORMS TO DIRECTIVES AND STANDARDS

Directives <i>Directives</i>	98/79/EC – IVD Medical Devices Classification : Outside Annex II and for self-testing Conformity Assessment Procedure : Annex III, section 6 2004/108/EC – Electromagnetic compatibility 2006/95/EC – Low voltage
Normes <i>Standards</i>	IEC 61010-1 : 2001 IEC 61010-2-081 : 2001 + A1 IEC 61010-2-101 : 2002 En 61326-1 : 2006 EN 61326-2-6 : 2006

Montpellier, France
19 août 2013
August 19th, 2013

Christian DUBUC

Directeur
Senior Director

TEMP-4018 Rev.9