

HORIBA ABX SAS
Parc Euromédecine
Rue du Caducée – BP7290
34184 MONTPELLIER Cedex 4 – France
Web : www.horiba.com
Tél. : 33 (0) 4 67 14 15 16

Déclaration de Conformité *Declaration of Conformity*

(No.dc10007h)

Selon
According to ISO 17050-1

NOUS LE FABRICANT *WE THE MANUFACTURER*

Nom <i>Name</i>	HORIBA ABX SAS
Adresse <i>Address</i>	Parc Euromédecine Rue du Caducée – BP7290 34184 MONTPELLIER Cedex 4 - FRANCE

DECLARONS QUE LE PRODUIT *DECLARE THAT THE PRODUCT*

Nom du produit <i>Product name</i>	Hematology analyzer
Modèles <i>Models</i>	ABX Pentra 120 – ABX Pentra 120 Retic – ABX Pentra 120 SPS – ABX Pentra 120 Retic SPS
Pays d'origine <i>Country of origin</i>	France

EST CONFORME AUX DIRECTIVES ET NORMES *CONFORMS TO DIRECTIVES AND STANDARDS*

Directives <i>Directives</i>	98/79/EC – IVD Medical Devices Classification : Outside Annex II and not for self-testing Conformity Assessment Procedure : Annex III 2004/108/EC – Electromagnetic compatibility
Normes <i>Standards</i>	IEC 1010-1 : 1990 + A1 : 1992 + A2 : 1995 EN 61326-1 : 2006 EN 61326-2-6 : 2006

Montpellier, France
30 septembre 2013
September 30th, 2013

Christian DUBUC
Directeur
Senior Director



TEMP-4018 Rev.9