

HORIBA ABX SAS
Parc Euromédecine
Rue du Caducée – BP7290
34184 MONTPELLIER Cedex 4 – France
Web : www.horiba.com
Tél. : 33 (0) 4 67 14 15 16

Déclaration de Conformité
Declaration of Conformity

(No.dc30004a)

Selon
According to ISO 17050-1

NOUS LE FABRICANT
WE THE MANUFACTURER

Nom <i>Name</i>	HORIBA ABX SAS
Adresse <i>Address</i>	Parc Euromédecine Rue du Caducée – BP7290 34184 MONTPELLIER Cedex 4 - FRANCE

DECLARONS QUE LE PRODUIT
DECLARE THAT THE PRODUCT

Nom du produit <i>Product name</i>	Clinical chemistry analyzer
Modèles <i>Models</i>	PENTRA C400 – PENTRA C400 Option I.S.E.
Pays d'origine <i>Country of origin</i>	France

EST CONFORME AUX DIRECTIVES ET NORMES
CONFORMS TO DIRECTIVES AND STANDARDS

Directives <i>Directives</i>	98/79/EC – IVD medical devices Classification : Outside Annex II and not for self-testing Conformity Assessment Procedure: Annex III 2004/108/EC – Electromagnetic compatibility
Normes <i>Standards</i>	IEC 61010-1 :2001 IEC 61010-2-081 : 2001 IEC 61010-2-101 : 2002 EN 61326-1 : 2006 EN 61326-2-6 : 2006 UL 61010-1 : 2004 CAN /CSA-C22.2 61010-1 : 2004

Montpellier, France
12 novembre 2013
November 12th, 2013



Christian DUBUC
Directeur
Senior Director

TEMP-4018 Rev.9