



HORIBA ABX SAS
Parc Euromédecine - Rue du Caducée
B.P. 7290
34184 Montpellier Cedex 4
FRANCE
www.horiba.com



LOT 1: 20633

04710802G



REF : A11A01650



2016/02/28

(FR) Constituant (EN) Constituent (DE) Bestandteil (ES) Componente (IT) Componente (PT) Componente (DA) Bestanddel (SV) Konstituent (EL) συστατικό (PL) Składnik (NO) Bestanddel	ABX Pentra 400	Pentra C400	Pentra C200	Unité Unit Einheit Unidad Unità Unidade Enhed Enhet Μονάδα Jednostka Enhet	Valeur titrée Assigned value Deklariert Wert Valor asignado Valori assegnati Valor determinado Tildelt værdi Tildelat värde Αναγραφόμενη τιμή Wartość przypisana Tildelt verd	Intervalle de confiance Confidence range Vertrauensintervalle Intervalos de confianza Livelli di fiducia Intervalo de confiança Konfidensområde Konfidensintervall Εύρος εμπιστοσύνης Zakres ufności Konfidensspekter	
C				U	$\oplus$	$\rightarrow \pm 3 \text{ SD} \leftarrow$	1 SD
CO ₂ (UV enzymatic)	●	●	●	mmol/L	20	16 - 24	1.3



HORIBA ABX SAS
Parc Euromédecine - Rue du Caducée
B.P. 7290
34184 Montpellier Cedex 4
FRANCE
www.horiba.com



LOT 1: 20633

04710802G



REF : A11A01650



2016/02/28

(FR) Constituant (EN) Constituent (DE) Bestandteil (ES) Componente (IT) Componente (PT) Componente (DA) Bestanddel (SV) Konstituent (EL) συστατικό (PL) Składnik (NO) Bestanddel	ABX Pentra 400	Pentra C400	Pentra C200	Unité Unit Einheit Unidad Unità Unidade Enhed Enhet Μονάδα Jednostka Enhet	Valeur titrée Assigned value Deklariert Wert Valor asignado Valori assegnati Valor determinado Tildelt værdi Tildelat värde Αναγραφόμενη τιμή Wartość przypisana Tildelt verd	Intervalle de confiance Confidence range Vertrauensintervalle Intervalos de confianza Livelli di fiducia Intervalo de confiança Konfidensområde Konfidensintervall Εύρος εμπιστοσύνης Zakres ufności Konfidensspekter	
C				U	$\oplus$	$\rightarrow \pm 3 \text{ SD} \leftarrow$	1 SD
CO ₂ (UV enzymatic)	●	●	●	mmol/L	20	16 - 24	1.3