

HORIBA ABX SAS  
Parc Euromédecine  
Rue du Caducée – BP7290  
34184 MONTPELLIER Cedex 4 – France  
Web : [www.horiba.com](http://www.horiba.com)  
Tél. : 33 (0) 4 67 14 15 16

## Déclaration de Conformité *Declaration of Conformity*

(No.dc10012e)

Selon  
According to ISO 17050-1

### NOUS LE FABRICANT *WE THE MANUFACTURER*

|                           |                                                                                    |
|---------------------------|------------------------------------------------------------------------------------|
| Nom<br><i>Name</i>        | <b>HORIBA ABX SAS</b>                                                              |
| Adresse<br><i>Address</i> | Parc Euromédecine<br>Rue du Caducée – BP7290<br>34184 MONTPELLIER Cedex 4 - FRANCE |

### DECLARONS QUE LE PRODUIT *DECLARE THAT THE PRODUCT*

|                                            |                  |
|--------------------------------------------|------------------|
| Nom du produit<br><i>Product name</i>      | <b>ABX e-Sat</b> |
| Modèles<br><i>Models</i>                   | -                |
| Pays d'origine<br><i>Country of origin</i> | France           |

### EST CONFORME AUX DIRECTIVES ET NORMES *CONFORMS TO DIRECTIVES AND STANDARDS*

|                                 |                                                                                                                                                                         |
|---------------------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Directives<br><i>Directives</i> | 98/79/EC – IVD Medical Devices<br>Classification : Outside Annex II and not for self-testing<br>Conformity Assessment Procedure : Annex III<br>2006/95/EC – Low voltage |
| Normes<br><i>Standards</i>      | IEC 61010-1 : 2010<br>EN 61326 : 1997 +A1 : 1998 +A2 : 2001 +A3 : 2003                                                                                                  |

Montpellier, France  
14 novembre 2013  
November 14th, 2013



**Christian DUBUC**  
Directeur  
Senior Director

TEMP-4018 Rev.9