

Hematology Devices (for in vitro diagnostic use)

ABX Minilyse LMG

Exclusive use:

Micros
Micros CRP
Pentra 60
Pentra 60 C+
Pentra 80
Pentra XL 80
Pentra 120
Pentra 120 Retic
Pentra DX 120
Slide Preparation System

15/01/05
A95A00010B

REF 0702010 (1L)

REAGENT 1L

IVD 



HORIBA ABX
BP 7290 - 34184 Montpellier
cedex 4 - France

1. Functions

Erythrocyte lysing reagent for white blood cell counting and differentiation, and hemoglobin determination on HORIBA ABX blood cell counters ^a.

Measurement procedure to be followed in using the device:

Principle of the method, specific analytical performance characteristics, analytical sensitivity, diagnostic sensitivity, analytical specificity, diagnostic specificity, accuracy, repeatability, reproducibility, limits of detection, limitations of the method and information about the use of available reference measurement procedures and materials by the user: see «Section: Specifications» in the instrument User Manual.

2. Conservation & expiration

Storage conditions: stored at 18°C (65°F) to 25°C (77°F). Product will degrade if exposed to air, keep cap / probe assembly, securely tightened.

Expiration date: refer to «expiration date» reagent packaging label.

3. Measurements, principles & results

Directions for use: see «Section: Reagent Location and connection» in the instrument User Manual. This reagent is for professional *in-vitro* diagnostic use only.

Measuring Principles : see «Section: Technology» in the instrument User Manual.

Results: see «Section: Results» in the instrument User Manual.

Performance data: see «Section: Specifications» in the instrument User Manual.

Note: if performance changes, call your HORIBA ABX representative.

4. Composition & handling precautions

Composition:

Potassium Cyanide..... < 0.1 %
Quaternary Ammonium Salt < 20%

pH: 10 +/- 0.5 (T=20°C)

Resistivity: 213 +/- 10 Ω (T=20°C)

Description: aqueous solution, limpid.

Handling Precautions: Avoid contact with eyes, skin and clothing. Wear laboratory gloves when handling the product. The product may be harmful if ingested. The product can be absorbed through an open wound, or inhalation. Please refer to the MSDS associated with the reagent.

Special precautions: Avoid contact with acid and aqueous acid environment: extremely toxic cyanide acid vapour can be formed. Please refer to the MSDS associated with the reagent.

Specimen Collection and Mixing: see «Section: Specimen collection and Mixing» in the instrument User Manual

Limitations: see «Section: Specifications» in the instrument User Manual

Safe Waste Disposal: Follow your laboratory's protocol when neutralizing and disposing of waste. Please refer to the MSDS associated with the reagent

