

HORIBA ABX SAS
Parc Euromédecine
Rue du Caducée – BP7290
34184 MONTPELLIER Cedex 4 – France
Web : www.horiba.com
Tél. : 33 (0) 4 67 14 15 16

Déclaration de Conformité *Declaration of Conformity*

(No.dc20025m)

Selon
According to ISO 17050-1

NOUS LE FABRICANT *WE THE MANUFACTURER*

Nom <i>Name</i>	HORIBA ABX SAS
Adresse <i>Address</i>	Parc Euromédecine Rue du Caducée – BP7290 34184 MONTPELLIER Cedex 4 - FRANCE

DECLARONS QUE LE PRODUIT *DECLARE THAT THE PRODUCT*

Nom du produit <i>Product name</i>	ABX CLEANER (1L) For application on : ABX Pentra 60, ABX Pentra 60 C+, Pentra ES 60, Pentra MS 60, Pentra MS CRP, ABX Pentra 80, ABX Pentra XL 80, Pentra XLR, ABX Pentra 120, ABX Pentra 120 Retic, ABX Pentra DX 120, ABX Pentra DF 120, Pentra DX Nexus, Pentra DF Nexus, Yumizen H500 OT ABX CLEANER (0.5L) For application on : ABX Micros ES 60, ABX Micros CRP, ABX Micros CRP 200,
Modèles <i>Models</i>	0903010 0903011
Pays d'origine <i>Country of origin</i>	France

EST CONFORME AUX DIRECTIVES ET NORMES *CONFORMS TO DIRECTIVES AND STANDARDS*

Directives <i>Directives</i>	98/79/EC – IVD Medical Devices Classification : General IVD (others) - Outside Annex II and not for self-testing Conformity Assessment Procedure: Annex III / Non-Annex II
Normes <i>Standards</i>	N/A

Montpellier, France
03 Février 2016
February 03rd, 2016


Christian DUBUC
Directeur
Senior Director

TEMP-4018 Rev.9