

# MINOCAL™ CALIBRATOR

## Hematology Reference Calibrator

Instrument:  
Performed by:  
Date:

Lot No.: CX407  
Exp. Date: 2017-03-05 (Y / M / D)

Table of Assigned Values	WBC K/ $\mu$ L	RBC M/ $\mu$ L	HGB g/dL	HCT %	PLT K/ $\mu$ L	MPV fL
ABX MICROS 45, ABX MICROS ABC VET	9.2 ± 0.2	4.57 ± 0.08	13.6 ± 0.2	38.8 ± 1.0	252 ± 12	8.8 ± 0.5
ABX MICROS 60, ADVIA® 60	9.4 ± 0.2	4.57 ± 0.08	13.7 ± 0.2	38.7 ± 1.0	246 ± 12	8.0 ± 0.5
ABX MICROS CRP 200	9.3 ± 0.2	4.58 ± 0.08	13.8 ± 0.2	39.0 ± 1.0	245 ± 12	8.7 ± 0.5
ABX PENTRA 60 OT, ABX PENTRA 60 C+	9.4 ± 0.2	4.58 ± 0.08	13.6 ± 0.2	36.9 ± 1.0	248 ± 12	9.7 ± 0.5
ABX PENTRA 80, ABX PENTRA XL80	9.3 ± 0.2	4.58 ± 0.08	13.7 ± 0.2	37.3 ± 1.0	250 ± 12	9.7 ± 0.5
ABX PENTRA XLR	9.3 ± 0.2	4.58 ± 0.08	13.7 ± 0.2	37.0 ± 1.0	250 ± 12	10.2 ± 0.5
ABX MICROS ES 60	9.2 ± 0.2	4.56 ± 0.08	13.6 ± 0.2	38.4 ± 1.0	243 ± 12	8.5 ± 0.5
ABX PENTRA DX 120, ABX PENTRA DF 120	9.5 ± 0.2	4.61 ± 0.06	13.6 ± 0.2	37.6 ± 1.0	252 ± 10	N/A
ABX PENTRA DX NEXUS, ABX PENTRA DF NEXUS	9.5 ± 0.2	4.61 ± 0.06	13.6 ± 0.2	37.6 ± 1.0	252 ± 10	N/A

Table 1. Determine Mean Value	WBC K/ $\mu$ L	RBC M/ $\mu$ L	HGB g/dL	HCT %	PLT K/ $\mu$ L
Sample 1					
2					
3					
4					
5					
6					
Total					
Mean Value (Total ÷ 6)					
SD					
%CV (SD ÷ Mean Value × 100)					

Table 2. Determine Calibration Factor	WBC K/ $\mu$ L	RBC M/ $\mu$ L	HGB g/dL	HCT %	PLT K/ $\mu$ L
MINOCAL Assigned Value Mean Value	—	—	—	—	—
Calibration Factor (round off to 3 decimal places)					
MINOCAL Reading (Table 1, Sample 6)					
Corrected Calibrator Reading for Adjustment					

Table 3. Verify Calibration	WBC K/ $\mu$ L	RBC M/ $\mu$ L	HGB g/dL	HCT %	PLT K/ $\mu$ L
Sample 1					
2					
3					
Total					
Mean Value (Total ÷ 3)					

For technical assistance, call the Customer Support Center at 1-888-903-5001.