

ABX Pentra

CO2 Control



HORIBA ABX SAS
Parc Euromédecine - Rue du Caducée
B.P. 7290
34184 Montpellier Cedex 4
FRANCE
www.horiba.com



LOT 1: 23211

REF : A11A01650

04710802H
SAP 1300023125



2018/01/31

(FR) Constituant (EN) Constituent (DE) Bestandteil (ES) Componente (IT) Componente (PT) Componente (DA) Bestanddel (SV) Konstituent (EL) συστατικό (PL) Składnik (NO) Bestanddel	ABX Pentra 400	Pentra C400	Pentra C200	Unité Unit Einheit Unidad Unità Unidade Enhed Enhet Μονάδα Jednostka Enhet	Valeur titrée Assigned value Deklariertes Wert Valor asignado Valori assegnati Valor determinado Tildelt værdi Tilldelat värde Αναγραφόμενη τιμή Wartość przypisana Tildelt verd	Intervalle de confiance Confidence range Vertrauensintervalle Intervalos de confianza Livelli di fiducia Intervalo de confiança Konfidensområde Konfidenstervall Εύρος εμπιστοσύνης Zakres ufności Konfidensspekter	
C				U	$\oplus$	$\rightarrow \pm 3 SD \leftarrow$	1 SD
CO2 (UV enzymatic)	•	•	•	mmol/L	20	16 - 24	1.3

ABX Pentra

CO2 Control



HORIBA ABX SAS
Parc Euromédecine - Rue du Caducée
B.P. 7290
34184 Montpellier Cedex 4
FRANCE
www.horiba.com



LOT 1: 23211

REF : A11A01650

04710802H
SAP 1300023125



2018/01/31

(FR) Constituant (EN) Constituent (DE) Bestandteil (ES) Componente (IT) Componente (PT) Componente (DA) Bestanddel (SV) Konstituent (EL) συστατικό (PL) Składnik (NO) Bestanddel	ABX Pentra 400	Pentra C400	Pentra C200	Unité Unit Einheit Unidad Unità Unidade Enhed Enhet Μονάδα Jednostka Enhet	Valeur titrée Assigned value Deklariertes Wert Valor asignado Valori assegnati Valor determinado Tildelt værdi Tilldelat värde Αναγραφόμενη τιμή Wartość przypisana Tildelt verd	Intervalle de confiance Confidence range Vertrauensintervalle Intervalos de confianza Livelli di fiducia Intervalo de confiança Konfidensområde Konfidenstervall Εύρος εμπιστοσύνης Zakres ufności Konfidensspekter	
C				U	$\oplus$	$\rightarrow \pm 3 SD \leftarrow$	1 SD
CO2 (UV enzymatic)	•	•	•	mmol/L	20	16 - 24	1.3