

HORIBA ABX SAS
Parc Euromédecine
Rue du Caducée – BP7290
34184 MONTPELLIER Cedex 4 – France
Web : www.horiba.com
Tél. : 33 (0) 4 67 14 15 16

Déclaration de Conformité *Declaration of Conformity*

(No.dc20006m)

Selon
According to ISO 17050-1

NOUS LE FABRICANT *WE THE MANUFACTURER*

Nom <i>Name</i>	HORIBA ABX SAS
Adresse <i>Address</i>	Parc Euromédecine Rue du Caducée – BP7290 34184 MONTPELLIER Cedex 4 - FRANCE

DECLARONS QUE LE PRODUIT *DECLARE THAT THE PRODUCT*

Nom du produit <i>Product name</i>	ABX DILUENT (20L) / ABX DILUENT (10L) For application on : ABX Pentra 60, ABX Pentra 60C+, Pentra ES60, Pentra MS60, Pentra MS CRP, ABX Pentra 80, ABX Pentra XL80, Pentra XLR, ABX Pentra 120, ABX Pentra 120 Retic, ABX Pentra DX120, ABX Pentra DF120, Pentra DX Nexus, Pentra DF Nexus, Yumizen H500 OT, Yumizen H500 CT, Yumizen H1500, Yumizen H2500
Modèles <i>Models</i>	0901020/0901010
Pays d'origine <i>Country of origin</i>	France

EST CONFORME AUX DIRECTIVES ET NORMES *CONFORMS TO DIRECTIVES AND STANDARDS*

Directives <i>Directives</i>	98/79/EC – IVD Medical Devices Classification : General IVD (others) - Outside Annex II and not for self-testing Conformity Assessment Procedure: Annex III / Non-Annex II
Normes <i>Standards</i>	N/A

Montpellier, France
05 Avril 2017
April 05th, 2017

Christian DUBUC
Directeur
Senior Director

TEMP-4018 Rev.9